



MicroBio Serial Number	010815140	PASS
Certificate Number	CC010815140 - 3	
		



MicroBio - Factory Calibration & Inspection Certificate

General Information

Product Owner	Majaseen LLC	Calibration Date	21 May 2019	Approved Signatory  S Plumridge Engineering Director	
Address	Ravila 14a-2019	Calibration Due	20 May 2020		
	Tartu	Calibrated By	SP		
	50411	Distributor		Customer PO/ Contact Name	Jane Oja
Country	Estonia	Calibration Methods	P0001W002 P0001W008*	CO/RMA Number	RMA-00245
				Equipment Asset Number	

Model / Configuration

Model	MB2-HiFlow	
Head	A/A 400 x 1.0mm (HiFlow Only)	
Plate Fitted	90 mm Petri dish	
Hardware/Firmware Ver.	2015	2.18
Total Volume (if known)	206594	
Total Samples (if known)	945	

Calibration Conditions

Parameter	Reference Conditions	Units
Allowed deviation	5% ±	
T lab temperature	294.45	K
ρ lab air pressure	1013.2	mbar
T_n standard temperature	293.15	K
ρ_n standard air pressure	1013.25	mbar
K_n compensation factor	0.9955	$\rho T_n / \rho_n T$

Visual inspection

Sampling head	OK
Battery box lid	OK
Battery contact springs	OK
Keypad surface / buttons	OK
Plate springs	OK
Head spring	OK
Enclosure	OK

AV2 Calibration System Details / Flow Conversion [Note 4]

Parameter	Value	Units
AV2 100 L/min set-point	2.07	m/s
AV2 180 L/min set-point	3.8	m/s
Q_s	180.00	L.min ⁻¹
V_a (AV2 @ Q_s)	3.8	m/s
K_{AV2}	47.37	L / ms.min
T	20.00	°C
ρ_n	1013.25	mbar
Serial Number	71341	
Calibration Date	22 Oct 2018	
Calibration Due Date	22 Oct 2019	
Certificate Number	CR2000	(MPB Ltd)
Accuracy	2% ±	

Measurements

Laboratory Temperature (T) [Note 3]	21.3	°C
Air pressure (ρ) [Note 2]	1013.2	mbar
Parameter	Mean Air Velocity (m.s ⁻¹)	Volumetric Flow (L.min ⁻¹)
Target	3.800	
Inspection Test	4.010	189.9
After adjustment	3.850	182.4

Notes / Recommendations

--

Calibration Results (95% Confidence Interval)

	Inspection	After Adjustment
Measured	189.9 L.min ⁻¹	182.37 L.min ⁻¹
Error	5.24%	1.30%
Result	ADJUST	PASS

Note 1: P0001W002 amended by P0001W008 - See <https://www.cantiumscientific.com/support/mb1-and-mb2-calibration-procedure/>. (Available to registered service agents only or upon request).
 Note 2: Greisinger GPB 3300 Digital Barometer
 Note 3: Greisinger GTH 1170 Digital Precision Quick Response Thermometer + Type-K Thermocouple probe
 Note 4: Airflow Development Limited AV2 air velocity meter + Cantium Scientific Limited air sampler calibration flow rig (see Note 1 - P0001W002)