


<b>MicroBio Serial Number</b>	<b>010815140</b>
<b>Certificate Number</b>	<b>CC010815140-6</b>

<b>New Sampler</b>	<input type="checkbox"/>
<b>Previous Certificate Suffix</b>	<b>5</b>

**General Information**

<b>Product Owner</b>	Majaseen LLC	<b>Calibration Date</b>	19 April 2022	<b>Approved Signatory</b>	 <small>S. Plumridge CEng MIET FRSA Director</small>
	Ravila 14a-2019	<b>Calibration Due</b>	19 April 2023		
<b>Address</b>	Tartu	<b>Calibrated By</b>	DGP	<b>PO / Name</b>	Jane Oja
	50411	<b>Distributor</b>		<b>CO / RMA</b>	RMA-00519
	Estonia	<b>Calibration Methods</b>	Procedure WI422r2	<b>Asset Number</b>	

**Model / Configuration**

<b>Model</b>	MicroBio MB2-HiFlow
<b>Sampler flow rate</b>	180 L/min
<b>Sampling head holes</b>	400 Holes
<b>Sampling head hole diameter</b>	1.0 mm
<b>Plate Fitted</b>	90mm Petri Dish
<b>Recorded Total Volume</b>	153306 Litres
<b>Recorded Total Samples</b>	743

**Sampler Calibration Requirements**

<b>Target Volumetric Flow</b>	180 L/min
<b>Target Anemometer Velocity</b>	4.99 m/s
<b>Allowed Deviation</b>	5.0 %
<b>Upper Anemometer Velocity</b>	5.2395 m/s
<b>Lower Anemometer Velocity</b>	4.7405 m/s

**Inspection**

<b>Sampling head</b>	OK
<b>Battery box lid</b>	OK
<b>Battery contacts</b>	OK
<b>Keypad</b>	OK
<b>Plate springs</b>	OK
<b>Head spring</b>	OK
<b>Enclosure</b>	OK
<b>Battery Pack</b>	Laboratory cells used

**qCR Kit Calibration Parameters**

Parameter	Value	Units
Calibration Factor	36.04	
Lab temperature	19.50	°C
Lab pressure	1015.4	mbar
Serial number	41426690101	
Calibration date	22 Jul 2021	
Calibration due date	22 Jul 2022	
Certificate number	CC41426690101-1	

**Measurements**

Reading No.	Inspection	Adjusted	
	Anemometer Target = 4.99 m/s		
1	5.19		m/s
2	5.09		m/s
3	5.11		m/s
4	5.14		m/s
5	5.17		m/s
6	5.14		m/s
7	5.13		m/s
8	5.10		m/s
9	5.07		m/s
10	5.12		m/s
11	5.10		m/s
12	5.02		m/s
<b>Mean Velocity</b>	5.115	0.000	m/s
<b>Std. Deviation</b>	0.0450	0.0000	
<b>Confidence 95%</b>	0.0255	0.0000	
<b>Upper Range</b>	5.1405	0.0000	m/s
<b>Lower Range</b>	5.0895	0.0000	m/s
<b>Action</b>	NONE	NONE	

**Sampler Performance**

	Inspection	Adjusted	
<b>Flow Rate Rate</b>	<b>184.34</b>		L/min
<b>Upper</b>	<b>185.26</b>		L/min
<b>Lower</b>	<b>183.43</b>		L/min
<b>d<sub>50</sub> (Theoretical)</b>	<b>1.67</b>		µm
<b>U<sub>j</sub> (Theoretical)</b>	<b>9.78</b>		m/s

**Notes****Inspection Modelling**

Q	# holes	Dj	Cc	ETA	stk50	Particle Density	d50 SQRT(Cc)	S <sub>50</sub>	d50	Uj	Aj
lpm		mm		g/cm-sec	0.22 circ	g/cm3	µm	mm	µm	m/sec	mm2
184.34	400	1.00	1.28	1.81E-04	0.22	1.03	1.89	0.110	1.67	9.78	0.785

**Adjustment Modelling**

Q	# holes	Dj	Cc	ETA	stk50	Particle Density	d50 SQRT(Cc)	S <sub>50</sub>	d50	Uj	Aj
lpm		mm		g/cm-sec	0.22 circ	g/cm3	µm	mm	µm	m/sec	mm2
	400	1.00	1.28	1.81E-04	0.22	1.03	0.00	0.110	0.00	0.00	0.785

Lab conditions: 20°C ±5% 1013.25mBar ±2.5%

 Calibration procedure: <https://www.dropbox.com/s/xbmqagajcak1lfh/WI422r2%20-%20MicroBio%20Calibration%20with%20Qualisair%20qCR%20Kit.pdf?dl=0>